

**Draft for your recall
information**

Your reference

Our reference

Tel. +49/56 61/71-

Fax +49/56 61/71-

Date

November 18, 2013

Recall of batches

Product: Original-Perfusor® Syringe 50 ml without cannula
REF: 8728844F
LOT: 13I2682029, 13I1482021, 13I0482026, 13I0482024, 13I0282026,
13H2282024, 13H1382027, 13H0782023

Dear,

In the course of internal quality checks we have found that with a low percentage the primary packaging (blister) of 'Original Perfusor® Syringes 50 ml' (version without Needle) may have 1-2 small holes on one of the edges of the transparent foil.

We would like to inform you, that we have decided to recall the above mentioned batches as a precautionary measure.

Your hospital has received **Original-Perfusor® Syringe 50 ml without cannula, REF: 8728844F** from the affected batches.

We would like to ask you not to continue with the use of syringes of the affected batches and return them to B.Braun. In parallel to this information our sales representatives are going to contact you for logistical support. A replacement has already been initiated.

Please forward this information to the users within your hospital immediately.

Please accept our apologies for any inconveniences.

We kindly ask you to confirm the receipt of this information.

Should you have any questions, please contact (name) at (telephone) or by e-mail@bbraun.com

Chairman of Supervisory Board:
Prof. Dr. h.c. Ludwig Georg Braun

Executive Board:
Dr. rer. pol. Heinz-Walter Große
(Vorsitzender)
Dr. rer. pol. Annette Beller
(Deputy member of the board)

Otto Philipp Braun
(Deputy member of the board)
Dr. rer. nat. Wolfgang Feller
Prof. Dr. med. Hanns-Peter Knaebel
Dr. rer. nat. Meinrad Lugan
Caroll H. Neubauer, LL.M.

Corporate Office: Melsungen
Reg. Gericht: Amtsgericht Fritzlar
HRB 11 000
WEEE-Reg.-Nr. DE 42690900

Address:
B. Braun Melsungen AG
Carl-Braun-Straße 1
34212 Melsungen
Germany

...

B | BRAUN

Page 2 to the letter of November 18, 2013 to

Yours sincerely,

.....