

**Aesculap AG  
Quality Management**

Postfach 40  
78501 Tuttlingen  
Germany

Contact: Del Solar  
Fon: [REDACTED]  
Fax: 07461 95-1555  
Email: [REDACTED]  
Internet: <http://www.aesculap.de>  
Date: August 10, 2017

**Safety Notice – Product Recall**

**NS357R – VEGA PS TIBIA TRIAL/PREPAR.PLATEAU T4**

-----  
You are using in your facility the instruments of the Aesculap AG Vega System®.

During the production process of one production unit, the Tibia Trial Plateau was laser marked with an incorrect size.

**The sizes "T3+" were falsely labeled "T4".**

The following drawing shows the size differences of T3+ and T4. If you lay them on top of each other, you will see the difference.

If they are of the same size, please send the Plateau labeled "T4" to the address below, we will send you the correct replacement immediately.


Chairman of Supervisory Board:  
Prof. Dr. h.c. Ludwig Georg Braun

Executive Board:  
Dr. Joachim Schulz  
(Chairman)  
Dr. Jens von Lackum

Corporate Office: Tuttlingen  
Register Court: Stuttgart HRB 726261  
VAT reg. no. DE812160059  
WEEE-Reg.-No. DE 65109852

Bank Account:  
**Deutsche Bank AG Tuttlingen**  
BLZ 653 700 75 Konto 21 22 000 00  
IBAN DE44 6537 0075 0212 2000 00  
SWIFT / BIC DEUTDE33  
**Baden-Württembergische Bank**  
BLZ 600 501 01 Konto 487 1905  
IBAN DE31 6005 0101 0004 8719 05  
SWIFT / BIC SOLADEST

Address:  
Aesculap AG  
Am Aesculap-Platz  
78532 Tuttlingen  
Germany



*Figure 1: Drawing of Tibia Trial Plateaus T3+ and T4*

In the case you do not have any of the affected products, please send us the attached **"Feedback Form"** and tick as appropriate.  
Should you have an affected product, please return it with the attached **"Product Recall Form"** to

**Local contact**

For any product-related request, kindly do not hesitate to contact our product manager

**Local contact**

We apologize for any inconvenience this may cause and thank you very much for your support.

With best regards,

Aesculap AG

[Redacted signature area]

Jasmin Francisca del Solar  
Quality Management Vigilance

**Feedback Form / FSCA**  
**Vega System Tibia Trial Plateau NS357R**

Please return this form by Fax or E-Mail to:

**Jasmin del Solar/ Department QMV**

**Fax +49 7461-95 1555**

**vigilance\_aag.de@aesculap.de**

We do not have any of the affected products in application

HOSPITAL \_\_\_\_\_ CITY \_\_\_\_\_

NAME \_\_\_\_\_ DEPARTMENT \_\_\_\_\_ FON \_\_\_\_\_

SIGNATURE \_\_\_\_\_

DATE \_\_\_\_\_



## PRODUCT RECALL

Hygienic condition:

new good

used decontaminated

used not decontaminated

pos. no.	part no. article no.	serial / lot-no.	quantity	remark	expectation	
					replacement of product	credit note
					<input type="checkbox"/>	<input type="checkbox"/>
					<input type="checkbox"/>	<input type="checkbox"/>
					<input type="checkbox"/>	<input type="checkbox"/>
					<input type="checkbox"/>	<input type="checkbox"/>
					<input type="checkbox"/>	<input type="checkbox"/>

### RETURN ADDRESS :

Aesculap AG  
LRP  
Siegfried Schwarz  
Am Aesculap-Platz  
D-78532 Tuttlingen – Germany

### ADDRESS / SENDER:

### DATE / SIGNATURE :

**Chairman of Supervisory Board:**  
Prof. Dr. h.c. Ludwig Georg Braun

**Executive Board:**  
Dr. Joachim Schulz  
(Chairman)  
Dr. Jens von Lackum

**Corporate Office: Tuttlingen**  
Register Court: Stuttgart HRB 726261  
VAT reg. no. DE812160059  
WEEE-Reg.-No. DE 65109852

**Bank Account:**  
**Deutsche Bank AG Tuttlingen**  
BLZ 653 700 75 Konto 21 22 000 00  
IBAN DE44 6537 0075 0212 2000 00  
SWIFT / BIC DEUTDE33  
**Baden-Württembergische Bank**  
BLZ 600 501 01 Konto 487 1905  
IBAN DE31 6005 0101 0004 8719 05  
SWIFT / BIC SOLADEV33

**Address:**  
Aesculap AG  
Am Aesculap-Platz  
78532 Tuttlingen  
Germany