



[Addressee name, address]

Date

Urgent Voluntary Field Safety Notice

Reference: R537

Purpose

This Field Safety Notice (FSN) is to inform you about a labeling issue of the outer packaging of the Arthrex Knee Scorpion™, AR-12990.

The Suture Passing Instruments product family consists of suture passers designed to grasp, deliver suture stitching, and perform suture retrieval in a single efficient step during an arthroscopic procedure.

Products affected by the issue

Product Name	Part No.	Lot No.	UDI
KNEE SCORPION™	AR-12990	15089076	00888867196322
		15089069	
		15095444	
		15095447	
		15095450	
		15095547	
		15096955	
		15096951	
		15096956	
		15103245	
		15104165	
		15104167	
		15112381	
		15112380	


Description of the issue

Associated outer labels and label content document for indicated batches display a sterility statement indicating “sterile” via irradiation but should indicate “nonsterile”.

DEFECTIVE



CORRECT (REFERENCE ONLY)



According to the Direction for use:

Devices provided non-sterile must be cleaned and sterilized prior to use or re-use; this is required as well for the first use after delivery of the unsterile instruments. Devices provided sterile must be cleaned and sterilized prior to each subsequent re-use (not for first use). Effective cleaning is an indispensable requirement for an effective sterilization of the instruments.

The change is recent for a product that has a long market history, so users may not recognize the change in label. Users should be familiar with the way the devices are packaged and are typically familiar with the devices before use and their status of being sterile or non-sterile.

The packaging is also not a peel pack and appears in a cardboard box and could be recognized as being unlikely to be sterile based on the packaging materials.

Knee Scorpions which are contained in loan sets are transferred to trays with other instruments and sterilized together before use.

All of this leads to the assessment for an improbable occurrence value.

Nevertheless, we would like to inform our customers that the label is incorrect and ensure that the device is reprocessed as described in the DFU.

To date, Arthrex is not aware of any adverse events associated with this issue.



Advise on action to be taken by the addressee of this notice

1. Immediately identify indicated product/batch numbers you have in your control.
2. Please make sure that the affected items are sterilized before used on patient as described in the DFU and discard the incorrect outer boxing.
3. Please complete the "Arthrex customer's response form" below and fax it back to +49 (89) 90 90 05 52 01 or email to vigilance@arthrex.de.
4. In case you have already experienced any issues with the related devices, please make sure to submit a complaint via email to complaints@arthrex.de.

Transmission of this Field Safety Notice

Please forward this Field Safety Notice (FSN) to all those who need to be aware of it within your organization or to any organization where the potentially affected devices have been transferred.

The relevant National Competent Authorities have been advised of this voluntary Field Safety Notice.

Contact information

If you have any questions, please call Arthrex GmbH at +49 89 90 90 05 52 40 and ask for Sarah Merkle. You can also send questions by email to complaints@arthrex.de.

Sincerely,

i.V. [REDACTED]

Manager Vigilance & Product Surveillance

Arthrex GmbH
Oskar-von-Miller-Str. 6
85235 Odelzhausen
Phone: +49 89 90 90 05 52 40
Fax: +49 89 90 90 05 52 01
Email: complaints@arthrex.de



Arthrex customer's response form

Field Safety Notice

Reference: R537

Return To		From	
To	Arthrex GmbH Product Surveillance Oskar-von-Miller-Str. 6 85235 Odelzhausen Germany	Facility Name	
Email	vigilance@arthrex.de	Address	
Fax	+49 89 90 90 05 52 01	City	
		Name	
		Title	

Please complete the form and return it by fax or email to the addressee above.

Hereby I confirm that we have been made aware of this Field Safety Notice that all affected devices must be sterilized prior to use and the outer boxing with the incorrect label must be discarded.

Date

Name

Signature