



Company: Nephros International Ltd.

Medical Device: OLpur MD 190

Attachment 5

Medical Device Product Withdrawal Notice

11 October 2005

Dear Customer: This notice includes information regarding the withdrawal of certain lots of OLpur MD 190 mentioned below. Following your receipt of this notice, a courier will be sent to pick up any remaining filters from the lot numbers listed below. Replacement product, if not already sent, will be shipped at that time. Please note that the competent authority will be notified of the decision to withdraw these lots.

MEDICAL DEVICE PRODUCT WITHDRAWAL NOTICE

Date Withdrawal Initiated: October 10, 2005

Withdrawal Initiated By: Manufacturer

Manufacturer: Nephros Inc.
3960 Broadway
New York, NY 10032
United States

Contact Person: Dara Gannon
Customer Service Manager (Nephros International Ltd)
+353 (0)1 213 5912

Product Trade Name: OLpur MD 190

Product Intended Use: Hemodiafiltration for patients with acute or chronic renal failure

Risk to Health: The use of, or exposure to, the product is not likely to cause any adverse health consequences

Reason for Withdrawal: Customer Satisfaction
(Possibility of fluid leaks during priming)

Products Withdrawn:	LOT	ESTIMATED QUANTITY TO BE WITHDRAWN (FOR ALL COUNTRIES)
	PI05/0254	792
	PI05/0311	768
	PI05/0319	705
	PI05/0320	1,380
	PI05/0339	1,470 (Germany)