


## MEDICAL DEVICE ADVISORY NOTICE


Information on Kehr Van Kemmel T Drain ref. GC6109, lot numbers 8093153, 8031593, 8064539 and 8074357

Dear Kehr Van Kemmel T Drain customer,


Coloplast informs you that for four lot numbers (8093153, 8031593, 8064539 and 8074357) there is a mislabelling on the box label: it is mentioned both "latex free" and "latex sleeve" on the same label (see picture below). The right information is only "LATEX SLEEVE". The product does contain latex.




~~IT-DRAINAGE /  
SILICONE / LATEX  
SLEEVE  
DRAIN ENT /  
SILICONE / GAINÉ  
LATEX~~



**Wrong labelling**



IT-DRAINAGE /  
SILICONE / LATEX  
SLEEVE  
DRAIN ENT /  
SILICONE / GAINÉ  
LATEX



**Correct labelling**

We advise customers to be aware not to use Van Kemmel T Drains for patients with known latex allergy.

Coloplast A/S is dedicated to patient safety thus this advisory notice has been made to inform customers that the Kehr Van Kemmel T Drain does contain latex. We sincerely apologise for any inconvenience this may have caused you.

Yours Sincerely,


  
Head of Regulatory Affairs



COLOPLAST  
 Centre d'affaires La Boursidière  
 92357 LE PLESSIS ROBINSON  
 France  
 Tel.: +33 (0)1 46 01 32 32



International Product Manager

  
 Fax.: +33 (0)1 46 01 32 64



COLOPLAST  
 Centre d'affaires La Boursidière  
 92357 LE PLESSIS ROBINSON  
 France  
 Tel.: +33 (0)1 46 01 32 32