

1499 Delp Drive

Harleysville, PA 19438

Tel: 215-256-4201

Fax: 215-256-1787

www.medcompnet.com

December 3, 2013

PRODUCT ALERT

Catalog Number: ASPC28-3PCE.

Lot No.: MFQB190


Dear Customer,

Our records indicate that you have received the above identified product. This product has been packaged with the incorrect catheter. The label indicates a 28cm Split Cath III catheter. The product was packaged with a 24cm catheter.

Medcomp is requesting the return of all affected product. Please contact your customer service representative for a Returned Goods Authorization (RGA) number.

We appreciate your assistance in this matter. Medcomp will investigate to determine the cause for the error, and take the appropriate steps to prevent a reoccurrence. If you have any questions or concerns please feel free to contact me directly.

Best regards,



Clinical Analyst / Regulatory
Medcomp.
(P) 215-256-4201 ext. [REDACTED]
(F) 215-256-9191
Email: [REDACTED]@medcompnet.com

cc: Regulatory File
Customer Service