



BIOMEDICA

Manufacture of medicinal products, medical devices, food supplements and cosmetics.
Purchase and sale of raw materials for the pharmaceutical, food and cosmetics.

URGENT FIELD SAFETY NOTICE/ DEVICE RECALL

Attention Distributors, Healthcare Professional

Details of affected devices

Commercial name of affected product: **NATRIUM CITRICUM 4%**
Batch. No. **9.178.01**

FSCA-Identifier: ID: LJRHV 1940, BfArM 17053/19
Intended use: Anticoagulation solution of trisodium citrate 4%
Type of Action: **Device Recall**, ID: LJSHVZP 2001

Description of the problem

Mould occurrence on the label on the primary packaging (bag) was noted. For the reason of patient safety decision about product recall was made.

Action to be taken by user

- Please stop using the device Natrium Citricum 4% 250 ml
- Please review your inventory (for batch no. 9.178.01)
- Return all unused identified products to manufacturer/distributors
- Share this notification with other relevant personnel in your organization

Transmission of this Field Safety Notice

This notice needs to be passed on the all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been supplied.


Contact reference person

Ing. Petra Honsů – Quality Manager Biomedica spol. s r.o.
Mob.: +420 770 105 577, e-mail: honsu.petra@bio-medica.eu

Ing. Stehlíková – Quality manager division Luhačovice, Biomedica spol. s r.o.
Mob.: +420 777 705 325, e-mail: stehlikova.dagmar@bio-medica.eu

Mrs. Magdalena Duhajská – Sales referent for infusion solutions, division Luhačovice, Biomedica spol. s r.o.
Mob.: +420 777 705 327, tel.: +420 57 713 10 27, e-mail: duhajska.magdalena@bio-medica.eu

Luhačovice 29.01.2020



Biomedica spol. s r.o.!