



# BIOMEDICA

Manufacture of medicinal products, medical devices, food supplements and cosmetics.  
Purchase and sale of raw materials for the pharmaceutical, food and cosmetics.

---

## URGENT FIELD SAFETY NOTICE/ DEVICE RECALL

**Attention** Distributors, Healthcare Professional

### Details of affected devices

Commercial name of affected product: **NATRIUM CITRICUM 4% 250 ml**  
Batch. No. **911540**

FSCA-Identifier: ID: LJRHV 2034, BfArM 21904/20  
Intended use: Anticoagulation solution of trisodium citrate 4%  
Type of Action: **Device Recall, ID: LJSHVZP 2101**

### Description of the problem

During plasma donation, an alarm is generated for the citrate tube because there is air in the tube. For the reason of patient safety decision about recall was made.

### Action to be taken by user

- Please stop using the device
- Please review your inventory,
- Return all unused identified products to manufacturer/distributors
- Share this notification with other relevant personnel in your organization

### Transmission of this Field Safety Notice

This notice needs to be passed on the all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been supplied.

### Contact reference person

Ing. Petra Honsů – Quality Manager Biomedica spol. s r.o.  
Mob.: +420 770 105 577, e-mail: [honsu.petra@bio-medica.eu](mailto:honsu.petra@bio-medica.eu)

Ing. Stehlíková – Quality manager site Luhačovice, Biomedica spol. s r.o.  
Mob.: +420 777 705 325, e-mail: [stehlikova.dagmar@bio-medica.eu](mailto:stehlikova.dagmar@bio-medica.eu)

Bc. Magdalena Duhajská – Sales referent for infusion solutions  
Mob.: +420 777 705 327, tel.: +420 57 713 10 27, e-mail: [duhajska.magdalena@bio-medica.eu](mailto:duhajska.magdalena@bio-medica.eu)

