

inomed Medizintechnik GmbH · Im Hausgrün 29 · D-79312 Emmendingen

To whom it may concern

Your contact: Saschka Busch
Department: Director QM/RA
Tel.: +49 7641 94 14 - 0
E-Mail: vigilance@inomed.com
Website: www.inomed.com

Date: 2022-11-22

Urgent Field Safety Notice – Voluntary Recall for disposable cannula 22Gx100x7

Dear valued customer,

with this notice we would like to inform you that inomed Medizintechnik GmbH is issuing a voluntary recall for article number **240114** disposable cannula 22Gx100x7 with batch number **22F074D** and **22K058D**.

This article has a nonconform coating.

We therefore advise you to discontinue the use of these cannula.
There are no reports regarding any adverse events or patient injuries available.
This recall is precautionary.

The risk for patients is considered low. Even under worst case circumstances it is very unlikely that the released amount of toxicologically relevant substances is high enough to cause adverse effects.

In order to support us in the best possible way, we kindly ask you to fill in the form on page 2. After returning the non-used items and confirming that the remaining stock has already been used, you will receive a corresponding credit note.

Best regards


inomed Medizintechnik GmbH
Saschka Busch 

Page 1 of 2

Inomed Medizintechnik GmbH
Im Hausgrün 29
79312 Emmendingen
Germany
Tel.: +49 7641 94 14 - 0
Fax: +49 7641 94 14 - 94
info@inomed.com
www.inomed.com

Authorised Representatives:
Dipl.-Ing. (FH) Rudi Mattmüller,
Mariusz Gruchala, Jörg Wipfler
Register place Freiburg
HRB 260788
VAT no. DE 141 993 531
Tax no. 05019/24280
WEEE-Reg. no. DE 55946849

National bank details:
Sparkasse Freiburg-
Nördlicher Breisgau
Account 2000 34 27
BIN: 680 501 01
BIC: FRSP DE 66
IBAN: DE 30 68050101
0020003427

International bank details:
Commerzbank Freiburg
Account 150 80 3500
BIN: 680 400 07
BIC: COBADEFF
IBAN: DE39 68040007
0150803500



Certified according to:
DIN EN ISO 13485

**Urgent Field Safety Notice
Voluntary Recall for
Disposable cannula 22Gx100x7**

Please complete this form and return it to vigilance@inomed.com.

If you have any issues in completing or return this form, please contact us as soon as possible.

Confirmation and disclosure form

I have read this Medical Device Field Safety Notice and understand its contents.
I will return all possible articles, which are affected of this Field Safety Notice.
We have already used the following articles from the affected batch number:

Already used articles from 22F074D: _____

Already used articles from 22K058D: _____

Name: _____

Position: _____

Healthcare facility / Distributer: _____

Email address: _____

Phone number: _____

Date: _____

Signature: _____