

Field Safety Note



Subject	Incorrect Product Labeling
Affected Product	LightCycler® Capillaries, 100 µl, M-Grade
Material Number	03612066001, lot 35374200
Required Actions	Immediately Discontinue Use of Affected Lot
Reference	RAS Product Removal Notice 07-Jul-2009
Date	08-Jul-2009
Author	██████████ RAS Quality Mangement

1. Issue

Roche Applied Science has confirmed that one lot of LightCycler® capillaries (100 µl, M-Grade, ID 03612066001, lot 35374200) has been incorrectly labeled. The label on the outer box indicates an expiry date of March 2010 instead of the correct expiry date of February 2009.

2. Potential Impact

The potential impact on diagnostic results is negligible as long as the product packaging is intact.

3. Corrections

The affected lot must not be used beyond February 2009.