



DIAsource ImmunoAssays[®] S.A.
Rue du Bosquet 2
1348 Louvain-La-Neuve
Belgium

Tel +32 10 84 99 11
Fax +32 10 84 99 90
www.diasource.be

FSCA Ref: CAPA31527
FSN Ref: CAPA31527

Date: 12 11 2019

Field Safety Notice

17 α -HYDROXYPROGESTERONE (17-OHP)-RIA-CT (#KIP1409)

To the attention of the users of the DIAsource 17-OHP RIA-CT assay - **lot 193606**

DIAsource ImmunoAssays Rue du Bosquet 2, B-1348 Louvain-la-Neuve + 32 10 84 99 11



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA31527
 FSN Ref: CAPA31527

Field Safety Notice (FSN)

17α-HYDROXYPROGESTERONE (17-OHP)-RIA-CT

Progesterone Tracer in the kits of 17OHP RIA

1. Information on Affected Devices*	
1.	1. Device Type(s)*
	Radioimmunoassay
1.	2. Commercial name(s)
	17α-HYDROXYPROGESTERONE(17-OHP)-RIA-CT
1.	3. Unique Device Identifier(s) (UDI-DI)
	NA
1.	4. Primary clinical purpose of device(s)*
	Radioimmunoassay for the in vitro quantitative measurement of human 17α-hydroxyprogesterone (17-OHP) in serum and plasma
1.	5. Device Model/Catalogue/part number(s)*
	KIP1409
1.	6. Software version
	NA
1.	7. Affected serial or lot number range
	193606 (expiry 15/11/2019)
1.	8. Associated devices
	NA

2 Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem*
	A very small proportion of the 17OH-Progesterone RIA kits of the lot 193606 might contain an erroneous tracer reagent corresponding to Progesterone (#3114583) instead of 17OH-Progesterone (#3114063). Both tracers having the same lot number 19H26/01, only the name and the article codes can distinguish them.
2.	2. Hazard giving rise to the FSCA*
	Using this erroneous tracer would result in abnormally low CPM for the Calibrators, controls and patients. Such profile of the calibration curve, as well as the B/B0 ratio (%)



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA31527
 FSN Ref: CAPA31527

	would make the run obviously invalid even if controls might give acceptable concentration values .
2.	3. Probability of problem arising Low
2.	4. Predicted risk to patient/users None
2.	5. Further information to help characterise the problem NA
2.	6. Background on Issue Coming from a user' complaint showing an erroneous tracer reagent corresponding to Progesterone (#3114583) instead of 17OH-Progesterone (#3114063)
2.	7. Other information relevant to FSCA NA

3. Type of Action to mitigate the risk*	
3.	<p>1. Action To Be Taken by the User*</p> <p> <input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input type="checkbox"/> Return Device <input checked="" type="checkbox"/> Destroy Device </p> <p> <input type="checkbox"/> On-site device modification/inspection </p> <p> <input checked="" type="checkbox"/> Follow patient management recommendations </p> <p> <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU) </p> <p> <input type="checkbox"/> Other <input type="checkbox"/> None </p> <p> Following the Notification sent by DIAsource on the 21/10/2019, we kindly asked you to check the contents of your 17OH Progesterone RIA kits of the lot 193606 in your inventory and let us know if any of them included the erroneous tracer. Now that the kits are close to expiry date (15/11/19), we assume that the content of your kits has been verified. We remind you that the kits which included a tracer of Progesterone, must not be used and have to be replaced. </p>



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA31527
 FSN Ref: CAPA31527

3.	2. By when should the action be completed?	15/11/2019
3.	3. Particular considerations for: IVD Is follow-up of patients or review of patients' previous results recommended? Yes We remind that the kits which included a tracer of Progesterone, by mistake, must not to be used and have to be replaced.	
3.	4. Is customer Reply Required? *	Yes (see reply's form)
3.	5. Action Being Taken by the Manufacturer <input type="checkbox"/> Product Removal <input checked="" type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input type="checkbox"/> Other <input type="checkbox"/> None Provide further details of the action(s) identified.	
3	6. By when should the action be completed?	15/11/2019
3.	7. Is the FSN required to be communicated to the patient /lay user?	No
3	8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet? Choose an item.	



DIAsource ImmunoAssays[®] S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA31527
 FSN Ref: CAPA31527

4. General Information*	
4.	1. FSN Type* New
4.	2. For updated FSN, reference number and date of previous FSN NA
4.	3. For Updated FSN, key new information as follows: NA
4.	4. Further advice or information already expected in follow-up FSN? * No
4	5. If follow-up FSN expected, what is the further advice expected to relate to: NA
4	6. Anticipated timescale for follow-up FSN NA
4.	7. Manufacturer information
	a. Company Name DIAsource ImmunoAssays S.A
	b. Address Rue du Bosquet 2, B-1348 Louvain-la-Neuve
	c. Website address www.diasource.be
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *
4.	9. List of attachments/appendices: NA
4.	10. Name/Signature [Redacted] [Redacted] [Redacted]

Transmission of this Field Safety Notice	
	This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)



DIAsource ImmunoAssays[®] S.A.
Rue du Bosquet 2
1348 Louvain-La-Neuve
Belgium
Tel +32 10 84 99 11
Fax +32 10 84 99 90
www.diasource.be

FSCA Ref: CAPA31527
FSN Ref: CAPA31527

<p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback..*</p>

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional