

It is of the utmost importance that the official in charge receives this information as soon as possible

Subject: Recall R003_20

Kremsmünster, 24. März 2020

Dear valued customer,

During the process of market surveillance, we have become aware that the below detailed MiniCollect® blood collection tubes (450409) partly show a wrong blood to additive ratio, which if used for blood sample analysis, might lead to falsification of analysis results.

For safety reasons, we have decided to recall the affected product.

We kindly ask you to **isolate** your inventory of this item and lot number on site and ensure that the affected lot is withdrawn from circulation at your customer's facilities.

Item Number	Lot Number	Product description
450409	191101	MiniCollect® TUBE 1 ml 9NC Coagulation sodium citrate 3.8% blue cap, with cross cut

Attached for your convenience are two templates for you to fill in, which will confirm to us that you have notified your customers and will also provide us with the total quantity that is isolated in your warehouse as well as any stock that is isolated in the warehouse of your customers.

Please complete the attached template and return it to our attention before March 30, 2020.

Once we have received the above requested information from you, we will provide you with further instructions.

Kindly accept our sincerest apologies for any inconvenience this matter may cause.

If you have any additional questions, please get in touch with your respective Greiner Bio-One Sales Representative or Assistant.

Best regards,

