

PRODUCT RECALL NOTIFICATION:**RE: anti-C4d Antibody, cat.no. BI-RC4D from lot# RCD191/RCD191C**

December 19, 2019

Dear valued Customer,

Our records indicate that you have received vials from the anti-C4d antibody, cat.no. BI-RC4D from lot# RCD191 and RCD191C.

If you are the end user of this product, please be aware of the following information. If you are not the end user, please forward a copy of this letter, along with the enclosed form to the end user.

We would like to inform you that the antibody from this specific anti-C4d antibody, lot# RCD191 and RCD191C, is not performing adequately and therefore Biomedica is issuing a recall of the lot in question. This specific anti-C4d lot# RCD191 and RCD191C was released in March 2019. You can trust your previous results if the positive control resulted in a positive staining.

If you have unused reagents in stock, please notify us of the number of unused vials in your inventory by attached form so that we send you a replacement. The unused vials should be destroyed on-site.

Please take our sincere apologies for the inconveniences this might have caused you.

Should you have any questions please contact Biomedica Medizinprodukte Customer Services at T +43-1-29107-45, F +43-1-29107-6389, or e-mail to: info@bmgrp.com.

Sincerely,



Certification of Product Destruction**Product Recall:** anti-C4d antibody (cat.no. BI-RC4D)

Recall file#: RC-BI-RC4D-191217

Antibody lots affected: #RCD191, #RCD191C

You have received vials of anti-C4d antibody from the above lot.

For our records please complete, sign and fax or email this document as proof of compliance to:**Customer Services - Biomedica Immunoassays****Biomedica Medizinprodukte GmbH**Fax: +43-1-29107-6389 or e-mail to info@bmgrp.com

| Anti-C4d Antibody Lot# | # Vials Used | # Vials on Stock |
|------------------------|--------------|------------------|
| RCD191/RCD191C | | |

Institution: _____

Name: _____

Title: _____

Signature: _____

Date: _____

