

WEINMANN Emergency Medical Technology GmbH + Co. KG  
PO Box 57 01 53 • 22770 Hamburg • GERMANY

Hamburg, July 2021

## **Important safety notice: Field safety corrective action on a medical device**

**Reference:** FSCA MCS2 2021-07.1

**From**

WEINMANN Emergency Medical Technology GmbH + Co. KG

**Addressee**

Users and operators, as well as specialist dealers and service partners

**Medical devices affected** (trade name and article no. of products):

- MEDUCORE Standard<sup>2</sup>; monitor/defibrillator WM 45300.
- Affected devices are listed in the annex.

Dear customers,

Quality and safety are our top priority, which is why we wish to act in a consistent and transparent manner as usual and, in the context of your obligation to co-operate under medical devices legislation, ask you to implement this corrective action so that users can continue to use our products on patients safely.

Page 1 of 4

**Company Headquarters**  
WEINMANN Emergency  
Medical Technology GmbH + Co. KG  
Frohbösestraße 12 • 22525 Hamburg • GERMANY  
T: +49 40 88 18 96-0  
F: +49 40 88 18 96-480  
www.weinmann-emergency.com

**Center for Production, Logistics, Service**  
WEINMANN Emergency  
Medical Technology GmbH + Co. KG  
Siebenstücken 14 • 24558 Henstedt-Ulzburg  
GERMANY

**Business Management**  
Dipl.-Volksw. Marc Griefahn  
Dipl.-Kfm. Philipp Schroeder  
Dipl.-Volksw. André Schulte

**Registration Court**  
Hamburg Municipal Court  
Dept. A # 115967  
V.A.T. # DE288367727  
WEEE Reg. # DE 47913245

**Creditor ID**  
DE35ZZZ00000353971

**General Partner**  
WEINMANN Emergency  
Management GmbH, Hamburg

**Registration Court**  
Hamburg Municipal Court  
Dept. B # 38144

**Certified QM System meeting**  
EC directive 93/42/EEC, Annex II  
(EN ISO 9001/EN ISO 13485)

**Banking Connections**

**Deutsche Bank AG Hamburg**  
IBAN DE87 2007 0000 0646 9639 00  
SWIFT DEUTDEHH

**Hamburger Sparkasse AG**  
IBAN DE44 2004 0000 0632 2626 67  
SWIFT HASPDEHHXXX

**Commerzbank AG Hamburg**  
IBAN DE14 2004 0000 0632 0071 00  
SWIFT COBADEHHXXX

### 1. Description of problem

In the course of our regular quality controls in Production, we noticed that during a particular production period (see annex: serial numbers affected), one of the electronic components of the MEDUCORE Standard<sup>2</sup> may deviate from original specifications during operation.

This can lead to a device malfunction (such as display of incorrect measured values) or even device failure.

### 2. Risk to the patient

The above fault may result in incorrect therapy, or even make therapy impossible.

### 3. Corrective action

The following corrective action must be performed:

- Check and, if necessary, replacement of the affected electronic component of the MEDUCORE Standard<sup>2</sup> by WEINMANN Emergency.

This corrective action is mandatory. The responsible authority has been informed of the procedure.

**You must not continue to use your MEDUCORE Standard<sup>2</sup>** until the corrective action described has been performed. You must decommission the device.

Please confirm receipt of this letter by **July 30<sup>th</sup>, 2021** and perform all actions by no later than **August 31<sup>st</sup>, 2021**.

The paragraphs below describe the process you need to follow.

a. **If you are an owner/operator, user or WEINMANN specialist dealer of MEDUCORE Standard<sup>2</sup> in Germany, proceed as follows:**

- Please use the attached report form to **confirm to us receipt of this letter, and that it has been forwarded as appropriate, by no later than July 30<sup>th</sup>, 2021.**
- Send all MEDUCORE Standard<sup>2</sup> devices to the address below to have the above-mentioned measures carried out:

**Center for Production, Logistics, Service  
WEINMANN Emergency Medical Technology GmbH + Co. KG  
Siebenstücken 14  
24558 Henstedt-Ulzburg, GERMANY**

- Please ensure in your organization that this **safety information is brought to the attention** of all users of the above-mentioned product and of other people to be informed.
- If you have passed these products on to third parties, **please forward a copy of this information to them or notify us of their contact information.**

b. **If you are an owner/operator or user of MEDUCORE Standard<sup>2</sup> in France, proceed as follows:**

- Please use the attached report form to **confirm to us that this letter has been received or forwarded, regardless of other actions**, by no later than **July 30<sup>th</sup>, 2021**.
- Please ensure in your organization that this **safety information is brought to the attention** of all users of the above-mentioned product and of other people to be informed.
- Contact the **WEINMANN service partner responsible for you**, referencing this Field Safety Notice (FSN) *FSCA MCS2 2021-07.1*, to coordinate and implement the above measures on all affected MEDUCORE Standard<sup>2</sup>.
- If you have passed these products on to third parties, **please forward a copy of this information to them or notify us of their contact information**.

c. **If you are an authorized WEINMANN service partner for MEDUCORE Standard<sup>2</sup> in France, proceed as follows:**

- Please use the attached report form to **confirm to us that this letter has been received or forwarded, regardless of other actions**, by no later than **July 30<sup>th</sup>, 2021**.
- Please ensure in your organization that this **safety information** is brought to the attention of all users of the above-mentioned products and of other people to be informed.
- Please ensure that this **safety information** is brought to the attention of all your customers for the above-mentioned products and of other people to be informed. **Please also pass this letter on to your customers for this purpose**.
- You will be receiving the document *Field Change Order FCO\_FSCA\_MCS2 2021-07.1* separately from this letter. The *Field Change Order* explains the procedure for performing the remedial measures described above.
- Please perform all **actions by no later than August 31<sup>st</sup>, 2021** and **confirm that the actions have been performed** on the document described in *Field Change Order*.

**Contact**

If you have any questions or need support, please contact your local specialist dealer or contact us directly:

Phone: +49 40 88 18 96 – 122

e-mail: [AfterSalesService@weinmann-emt.de](mailto:AfterSalesService@weinmann-emt.de).

Kind regards,

WEINMANN Emergency  
Medical Technology GmbH + Co. KG



André Schulte  
Managing Director



p.p. Dennis Horstmann  
Authorized Signatory  
Head of Supply Chain + Quality Management

**Annexes**

„Field safety notice received“ report form  
Table of serial numbers affected

# Report to WEINMANN Emergency by 2021-07-30

Regarding MEDUCORE Standard<sup>2</sup> safety information: Reference: FSCA MCS2 2021-07.01

Original letter sent to:

Please fill in this report form in full and return it by e-mail, fax or mail to:

e-mail: **AfterSalesService@weinmann-emt.de**  
Fax: **+49 40 88 18 96 - 490**

**WEINMANN Emergency Medical Technology GmbH + Co. KG**  
Technical Service  
Frohbösestraße 12  
22525 Hamburg, GERMANY

- I hereby confirm receipt of this letter and that I have read and understood its contents. This letter has been brought to the attention of all users of the product and of other people in my organization who need to be informed.  
If the products have been passed on to third parties (applies to specialist dealers, for example), a copy of this information has been passed on to them.

Please complete in full in block capitals:

- Company/organization details are identical to those of the addressee above.  
 Company/organization details differ from those of the addressee as follows:

Customer no.:

\_\_\_\_\_

Company/organization + address:

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

- I am no longer in possession of the medical device:

The device has been scrapped

The new owner is (company + address)

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_  
Date, signature

\_\_\_\_\_  
Name (in block letters)

\_\_\_\_\_  
Position (in block letters)

\_\_\_\_\_  
e-mail address (in block)

