

## URGENT Field Safety Notice

M5071A (adult) and M5072A (infant/child) pads for use with HS1/OnSite/Home AEDs may experience gel separation and reduction of gel surface area

21-FEB-2022

**This document contains important information for the continued safe and proper use of your equipment**

Please review the following information with all members of your staff who need to be aware of the contents of this communication. It is important to understand the implications of this communication.

Please retain a copy with the equipment Instruction for Use.

Dear Customer,

A problem has been identified in the Philips M5071A (adult) and M5072A (infant/child) AED pads that could pose a risk for patients or users. This URGENT Field Safety Notice is intended to inform you about:

### 1. What the problem is and under what circumstances it can occur

HS1/OnSite/Home AED pads (PN: M5071A, M5072A) have been observed to experience gel separation from the foam/tin backing when peeled from the yellow plastic liner. The gel may fold onto itself resulting in reduced surface area of gel on the pad, or it may separate almost completely leaving only a small amount of gel on the pad. Any pad currently installed in or stored with an HS1/OnSite/Home AED could experience this problem, and it is not possible to know prior to patient use if your pad is affected because the pads are protected by a foil seal. Philips has received 115 complaints about this issue since 2010 (of which 84 complaints were received in 2021) for a total of approximately 5 million shipments of M5071A and M5072A pads. Users should continue to use the HS1/OnSite/Home AED and pads as-is, and follow the voice prompts because the AED will step the user through the necessary actions.

The HS1/OnSite/Home AED is intended for use by minimally trained or untrained individuals (e.g., individual homeowners, institutional response team members, teachers, and coaches) to treat victims of suspected sudden cardiac arrest.

## 2. Describe the hazard/harm associated with the issue

When a pad with separated, folded gel is placed on the patient's bare skin, the HS1/OnSite/Home AED could deliver less effective or ineffective therapy to the patient due to the reduced surface contact area with the skin. See example picture in **Figure 1**.

Separated, folded gel may also have a discolored and/or melted appearance. While the gel may also have a discolored and/or melted appearance, the appearance does not have any impact on the delivery of therapy; however, there may be a delay in therapy if the user hesitates to apply the pad due to its appearance, and the AED may deliver less effective or ineffective therapy to the patient due to the reduced contact area with the skin. See example picture in **Figure 2**.

It is also possible that the gel could separate almost completely from the foam/tin backing when peeled, (see **Figure 3**.) Due to a small amount of gel surface contact area with the skin, electrical arcing could occur when a shock is delivered leading to burns to the patient's skin, or the AED could be unable to deliver any shock through the pads. A delay in therapy will result while the user installs a replacement pads cartridge (if available) or performs CPR while waiting for Emergency Medical Services Personnel to arrive. For comparison, **Figure 4** shows a normal pad. No matter the state of the pad, follow the voice prompts because the AED will step you through the necessary actions.




**Figure 1:** Separated gel that has folded onto itself when peeled.

Action:  
Apply pads to the patient.  
Do not hesitate.




**Figure 2:** Separated, folded gel may also have a discolored and/or melted appearance.

Action:  
Apply pads to the patient.  
Do not hesitate.



**Figure 3:** Gel almost completely separated from backing.

Action:  
Replace pads cartridge if a spare is available. If no spare is available, perform CPR until help arrives.



**Figure 4:** Normal pad.

Action:

Apply pads to the patient according to the Instructions for Use/Owner's Manual.

### 3. Affected products and how to identify them

Affected products include all Lots of Adult and Infant/Child Pads Cartridges (PNs: M5071A and M5072A) installed in or stored as spares with the HS1, OnSite, and Home AEDs. This notice takes into consideration only pads that are unexpired. Note, subsequent shipments will still be affected until updated pads are available.

The M5071A and M5072A part numbers are located on the pads cartridge and the foil packaging. The M5072A identifier can also be found on the box that Infant/Child pads are shipped in. See photos below with the location of the part number circled.



**M5071A**  
foil packaging



**M5071A**  
pads cartridge



**M5072A** box



**M5072A**  
foil packaging



**M5072A**  
pads cartridge

**4. Describe the actions that should be taken by the customer / user in order to prevent risks for patients or users**

Continue using the HS1/OnSite/Home AED and pads as-is. During use, ensure the majority of the pad surface is covered with gel and apply the pads to the patient. If you notice the gel beginning to separate from the foam backing as you peel, try to prevent the gel from folding onto itself if possible. Do not hesitate to apply the pads to the patient unless the gel has almost completely separated from the backing as in **Figure 3**. In case of trouble, install spare pads if available and continue the rescue. No matter the state of the pads, follow the voice prompts because the AED will step you through the necessary actions.

Do not try to examine the pads gel prior to patient use. It is not possible to know if your pads are affected by the problem prior to use because the pads are protected by a foil seal. The foil seal on the pads cartridge should be opened only for patient use in an emergency because the pads will quickly dry out if the foil seal is broken.

Philips recommends that you store a spare pads cartridge with your HS1/OnSite/Home AED. A short video showing how to replace the pads cartridge can be found at: [www.philips.com/replace-aed-pads-video](http://www.philips.com/replace-aed-pads-video)

If the problem continues and you do not have a spare pads cartridge, attend to the patient, providing CPR if needed, until Emergency Medical Services Personnel arrive.

Please pass this notice to all those who need to be aware within your organization or to any organization where HS1/Onsite/Home AED devices or pads cartridges have been transferred, (if appropriate.)

Keep a copy of this letter with the Instructions for Use/Owner's Manual of your HS1/OnSite/Home AED because subsequent shipments of M5071A and M5072A pads will still be affected until updated pads are available.

Finally, please complete and return the Reply Form found at the end of this letter.

**5. Describe the actions planned by the Philips Emergency Care and Resuscitation business to correct the problem**

Philips is actively working on design changes intended to eliminate the issue in the M5071A and M5072A pads. Philips projects to release updated pads later in 2022, dependent upon design activities, subcomponent availability, and regulatory approvals. Philips plans to notify eligible customers and supply updated pads.



## 6. Additional information

Users should follow the voice prompts because the AED will step you through the necessary actions. As described in the Instructions for Use, you may hear voice prompts to assist you as shown below.

HS1/OnSite/Home tells you:	Possible cause	Recommended action
<b>...to insert a pads cartridge</b>	The pads cartridge has been damaged.	Insert a new pads cartridge.
<b>...to press pads firmly to the skin</b> <b>...to make sure the pads have been removed from the liner</b> <b>...the pads should not be touching the patient's clothing.</b>	The pads are not properly applied to the patient.	Make sure the pads are sticking completely to the patient's skin.
<b>...to insert new pads cartridge</b>	The pads cartridge has been opened and the pads peeled off the liner, but the pads have not been successfully attached to the patient. There may be a problem with the pads cartridge.	Replace the damaged pads cartridge. Pull up the handle on the cartridge cover, and replace pads on patient with new pads to continue with the rescue.

If you need any further information or support concerning this issue, please contact your local Philips representative. [< Key Markets insert contact information here. >](#)

This notice has been reported to the appropriate Regulatory Agencies. Be sure to report any occurrence of this issue to Philips, your Philips representative, or to your local Regulatory authority.

Philips regrets any inconvenience caused by this problem.

Sincerely,



**URGENT FIELD SAFETY NOTICE RESPONSE FORM**  
**Reference: Gel Separation, M5071A and M5072A, 2021-CC-EC-012**

**Instructions:** Please complete and return this form promptly and no later than 30 days from receipt. Completing this form confirms receipt of the Urgent Field Safety Notice Letter, understanding of the issue, and required actions to be taken.

Customer/Consignee/Facility Name: \_\_\_\_\_

Street Address: \_\_\_\_\_

City/State/ZIP/Country: \_\_\_\_\_

**Customer Actions:**

- Continue using the pads as-is. During use, ensure the majority of the pad surface is covered with gel and apply the pads to the patient. If you notice the gel beginning to separate from the foam backing as you peel, try to prevent the gel from folding onto itself if possible. Do not hesitate to apply the pads to the patient unless the gel has almost completely separated from the backing. In case of trouble, install spare pads if available and continue the rescue. No matter the state of the pads, follow the voice prompts because the AED will step you through the necessary actions.
- Do not try to examine the pads gel prior to patient use. It is not possible to know if your pads are affected by the problem prior to use because the pads are protected by a foil seal. The foil seal on the pads cartridge should be opened only for patient use in an emergency because the pads will quickly dry out if the foil seal is broken.
- Consider storing a spare pads cartridge with your HS1/OnSite/Home AED. A short video showing how to replace the pads cartridge can be found at: [www.philips.com/replace-aed-pads-video](http://www.philips.com/replace-aed-pads-video)
- If the problem continues and you do not have a spare pads cartridge, attend to the patient, providing CPR if needed, until Emergency Medical Services Personnel arrive.
- Please pass this notice to all those who need to be aware within your organization or to any organization where HS1/Onsite/Home AED devices or pads cartridges have been transferred, (if appropriate.)
- Keep a copy of this letter with the Instructions for Use/Owner's Manual of your HS1/OnSite/Home AED.

I acknowledge receipt and understanding of the accompanying Urgent Field Safety Notice Letter and confirm that the information from this Letter has been properly passed to those who need to be aware.

**Name of person completing this form:**

Signature: \_\_\_\_\_ Date (DD/MM/YYYY): \_\_\_\_\_

Printed Name: \_\_\_\_\_ Telephone Number: \_\_\_\_\_

Title: \_\_\_\_\_ Email Address: \_\_\_\_\_

Please return this form to Philips by email or fax < [Key Market Insert contact information](#) >

## URGENT Field Safety Notice (Expansion)

M5071A Adult and M5072A Infant/child pads cartridges for use with HS1/OnSite/Home AEDs may experience gel separation and reduction of gel surface area

15-NOV-2022

**This document contains important information for the continued safe and proper use of your equipment**

Please review the following information with all members of your staff who need to be aware of the contents of this communication. It is important to understand the implications of this communication.  
Please retain a copy with the equipment Instructions for Use.

Dear Customer,

On March 4, 2022, Philips voluntarily initiated an "Urgent Field Safety Notice" for the M5071A Adult and M5072A Infant/child pads cartridges for use with HS1/OnSite/Home AEDs. Due to customer feedback, we have decided to notify customers who received an HS1/OnSite/Home AED more than 10 years ago. **Please respond, even if you no longer own the AED.** This notice has been reported to the appropriate Regulatory Agencies. This letter is intended to inform you about:

### 1. What the problem is and under what circumstances it can occur

Some electrode pads used with the HS1/OnSite/Home AED have been observed to experience gel separation from the foam/tin backing when peeled from the yellow plastic liner. The gel may fold onto itself, resulting in reduced surface area of gel on the pad. A pad in this condition could cause the HS1/OnSite/Home AED to deliver less effective or ineffective therapy to the patient due to the reduced surface contact area with the skin.



Areas of missing gel

Separated, folded gel may also have a discolored or melted appearance. While the gel may also have a discolored or melted appearance, the appearance does not have any impact on the delivery of therapy; however, there may be a delay in therapy if the user hesitates to apply the pad due to its appearance, and the HS1/OnSite/Home AED may deliver less effective or ineffective therapy to the patient due to the reduced contact area with the skin.



It is also possible that the gel could separate almost completely from the foam/tin backing when peeled. Due to a small amount of gel surface contact area with the skin, electrical arcing could occur when a shock is delivered leading to burns to the patient's skin, or the HS1/OnSite/Home AED could be unable to deliver any shock through the pads. A delay in therapy will result while the user installs a replacement pads cartridge (if available) or performs CPR while waiting for Emergency Medical Services Personnel to arrive.



Gel almost completely missing

**2. Hazard/harm associated with the issue:** An electrode pad that experiences gel separation could result in less effective therapy for a patient, delay of therapy for a patient, or cause the HS1/OnSite/Home AED to be unable to deliver any shock through the pads.

**3. Affected products and how to identify them:** M5071A Adult and M5072A Infant/child pads cartridges with a LOT number that begins with "Y" may experience the issue.

**LOT Number of affected pads begins with "Y"**



#### 4. Actions that should be taken by the customer/user in order to prevent risks for patients or users

Keep your HS1/OnSite/Home AED in service until you receive updated pads. If you need to use your HS1/OnSite/Home AED before an updated pads cartridge has been installed, ensure the majority of the pad surface is covered with gel and apply the pads to the patient. If you notice the gel beginning to separate from the white foam backing as you peel, try if possible to prevent the gel from folding onto itself. Do not hesitate to apply the pads to the patient unless the gel has almost completely separated from the backing. In case of trouble, install a spare pads cartridge if available and continue the rescue. No matter the state of the pads, follow the voice prompts because the HS1/OnSite/Home AED will guide you through the necessary actions. If the problem continues and you do not have a spare pads cartridge, attend to the patient, providing CPR if needed, until Emergency Medical Services Personnel arrive. Philips recommends that you store a spare pads cartridge with your HS1/OnSite/Home AED. A short video showing how to replace the pads cartridge can be found at: [www.philips.com/replace-aed-pads](http://www.philips.com/replace-aed-pads)

If your HS1/OnSite/Home AED is in service and installed with an unexpired M5071A Adult pads cartridge with a LOT number that begins with "Y", then you are eligible to receive an updated Adult pads cartridge, free-of-charge. Unexpired M5071A spare pads cartridges with a LOT number that begins with "Y" will be replaced, free-of-charge. You must respond to receive any free-of-charge updated Adult pads cartridges. Please pass this notice to all those who need to be aware within your organization or to any organization where HS1/OnSite/Home AED devices or pads cartridges have been transferred. Please keep a copy of this letter with the Instructions for Use/Owner's Manual. **Please respond to this notice because your response is necessary to ensure effectiveness of the recall notification.** Even if you transferred your device to someone, or if your device is no longer in service, please respond.


#### 5. Actions planned by Philips to correct the problem

Philips will provide to customers who respond one updated Adult pads cartridge, free-of-charge, per HS1/OnSite/Home AED in service that is installed with an unexpired M5071A Adult pads cartridge with a LOT number that begins with "Y". Unexpired M5071A spare pads cartridges with a LOT number that begins with "Y" will be replaced, free-of-charge. You must respond to receive any free-of-charge updated Adult pads cartridges.

Infant/child pads cartridge updates will be handled separately. If you own an M5072A Infant/child pads cartridge, Philips will provide, free-of-charge, updated M5072A Infant/child pads cartridges when available to replace unexpired Infant/child pads cartridges.

A similar notice was previously sent to customers who purchased HS1/OnSite/Home AEDs less than 10 years ago. **If you received that notification, please respond to both notifications.** If you need any further information or support concerning this issue, please contact your Philips representative. Philips regrets any inconvenience caused by this problem.

Sincerely,



Director of Quality



## Reply Form for FSN C&R 2022-CC-EC-011

**Instructions:** Whether your HS1/OnSite/Home AED remains in service or not, please respond within the next 30 days. < Add text and/or graphics to suit the method of response you want, e.g., QR code, URL, email, fax, etc. Example: "Please write your answers below, provide your signature and contact info, and return the form. You may take a photo of the completed form or scan with a scanner. Email or fax to: *Insert Philips local Key Market email address and fax number here.*" >

1. What is your name and the name of your facility?

Write your name and the name of your facility.


2. What is the status of your HS1/OnSite/Home AED?

Please check the box and use the space on the right to provide additional info. If different status for multiple AEDs please provide information for all AEDs and make the status of each clear in your response.


- My HS1/OnSite/Home AED remains in service, and my M5071A Adult pads cartridge is unexpired and has a LOT number that begins with "Y". The **Serial Number** of my HS1/OnSite/Home AED and the **LOT number(s)** of my M5071A Adult pads cartridge, and spare cartridge(s) if applicable, are listed to the right. (If multiple HS1/OnSite/Home AEDs and Adult pads, please include all. Add a separate page if needed.) Therefore, please send me the replacement Adult pads cartridge(s). I understand I need to provide both the serial number of the device and LOT number(s) of the unexpired Adult pads cartridge(s) to receive the replacement(s) free-of-charge. (Note: M5072A Infant/child pads cartridge updates will be handled separately, so please do not include Infant/child pads cartridge information.)
- My HS1/OnSite/Home AED is no longer in service (discarded, lost, permanently retired or destroyed). If the Serial Number is known to me, I have listed it to the right. (If multiple, please include all serial numbers.)
- Ownership of my HS1/OnSite/Home AED was transferred to someone else. If available, I have listed the Serial Number of the AED to the right. (If multiple, please include all serial numbers. Add a separate page if needed.) I will share this notice with the new owner(s).
- My HS1/OnSite/Home AED is installed with an Adult pads cartridge that is past its expiration date. I have listed the AED Serial Number(s) in the space provided, but I understand that because my pads are expired, I will not receive a free-of-charge replacement Adult pads cartridge.

Use this space to provide additional information.

Serial number example



LOT # example



I acknowledge receipt and understanding of this Urgent Field Safety Notice and confirm that the information from this notification has been properly distributed to all who handle the HS1/OnSite/Home AED. The contact information below will be used to update Philips' records in case of future safety notices, and to provide the replacement pads cartridge(s). The information below will not be used for marketing purposes. Acknowledging this notice will not reduce your coverage or rights under any Philips AED Warranty or Indemnification.

Signature: \_\_\_\_\_

Printed Name: \_\_\_\_\_

Address:

(Facility address or personal address, as applicable)

Telephone Number: \_\_\_\_\_

Email Address: \_\_\_\_\_

Date (DD-MMM-YYYY): \_\_\_\_\_