


Date: 11.08.2020

**Urgent Field Safety Notice**  
**Dispenser DP 30 LipoPlus**

For Attention of: Nouvag GmbH. – Germany

Contact details of local representative
<b>Nouvag AG</b> <b>Mehdi Zadehnour</b> <b>St. Gallerstrasse 23-25</b> <b>9403 Goldach</b> <b>+41 71 846 66 57</b>


**Urgent Field Safety Notice (FSN) DP 30 LipoPlus**  
**Production according to expired EMV Standard 60601-1-2 Edition 3**

1. Information on Affected Devices	
1.	<p>1. Device Type</p> <p>The Dispenser DP 30 LipoPlus is a specifically for liposuction designed tumescence infiltration pump, delivering high volume of tumescence liquid.</p> 
1.	<p>2. Commercial name(s)</p> <p>Dispenser DP 30 LipoPlus</p>
1.	<p>3. Unique Device Identifier(s) (UDI-DI)</p> <p>+ENOU41610F +ENOU41630H</p>
1.	<p>4. Primary clinical purpose of device(s)*</p> <p>The DP 30 LipoPlus is a mobile Infiltration pump that is used for Tumescence infiltration during Liposuction and for treatments in Angiology</p>
1.	<p>5. Device Model/Catalogue/part number(s)</p> <p>4161 and 4163</p>

1.	6. Affected serial or lot number range		
	Qty	SET SN	UNIT SN
	1	7438S1906R	4606U1904R
	1	3501S2004R	5872U1911R
	1	4435S2005R	5874U1911R
	1	5884E1905R	7914U1901R
	1	0627S1907R	4613U1904R
	1	0628S1907R	4614U1904R

2 Reason for Field Safety Corrective Action (FSCA)	
2.	1. Description of the product problem
	The devices DP 30 LipoPlus does not comply with the latest harmonized EMC standard (60601-1-2, Edition 4). The device only complies with the expired Edition 3 and was not adapted to the new standard.
2.	2. Hazard giving rise to the FSCA
	The device might interfere with other electrical devices. The DP 30 LipoPlus could disturb the function of devices nearby or could itself be disturbed by them.
2.	3. Probability of problem arising
	Little to no probability of problems arising. The device still complies with the previous Edition 3 EMC standard (IEC 60601-1-2:2007). With the harmonization of the EMC standard Edition 4 (IEC 60601-1-2:2014) the acceptable ranges of electromagnetic interference is now smaller and thus not successfully achieved by the device.
2.	4. Predicted risk to patient/users
	none

<b>3. Type of Action to mitigate the risk</b>		
<b>3.</b>	<b>1. Action To Be Taken by the User</b>  <input checked="" type="checkbox"/> Identify Device <input type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input type="checkbox"/> Destroy Device  <input type="checkbox"/> On-site device modification/inspection  <input type="checkbox"/> Follow patient management recommendations  <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU)  <input type="checkbox"/> Other <input type="checkbox"/> None	
	Device must be returned to the following address:  Nouvag GmbH Dental und Medizintechnik Schulthaissstrasse 15 DE - 78462 Konstanz Germany  Tel. +49 (0)7531 1290-0 Fax +49 (0)7531 1290-12 <a href="mailto:info-de@nouvag.com">info-de@nouvag.com</a>	
<b>3.</b>	<b>2. By when should the action be completed?</b>	Immediately
<b>3.</b>	<b>3. Is customer Reply Required? *</b> (If yes, form attached specifying deadline for return)	Yes, As soon as possible
<b>3.</b>	<b>4. Action Being Taken by the Manufacturer</b>  <input type="checkbox"/> Product Removal <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input checked="" type="checkbox"/> Other <input type="checkbox"/> None  Device modification on manufacturing site	

4. General Information	
4.	1. FSN Type New
4.	2. For updated FSN, reference number and date of previous FSN N/A
4.	3. For Updated FSN, key new information as follows: N/A
4.	4. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)
	a. Company Name <b>Nouvag AG</b>
	b. Address <b>St. Gallerstrasse 23-25, CH-9403 Goldach</b>
	c. Website address <b>www.nouvag.com</b>
4.	5. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.
4.	6. Name/Signature Mehdi Zadehnour, COO 

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p>



It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.

**Please fill in the customer/ distributor reply form and send it to us before the defined deadline at: [vigilance@nouvag.com](mailto:vigilance@nouvag.com)**