



Important Medical Device Recall Notice

July [XX], 2006

RE: Recall Focus® DAILIES® Toric

Dear Eye Care Professional,

[Following our recent telephone conversation we are writing to confirm that] / (We are writing to inform you that) CIBA Vision is conducting a recall on a limited number of lots of Focus DAILIES Toric disposable soft contact lenses. A small amount of lenses in these lots may be labelled with the incorrect power. There is no health risk involved There is no health risk involved with the usage of an incorrectly labelled lens because consumers are likely to immediately be aware, upon insertion, that such a lens has the wrong power.

Our records indicate that you may have received lenses from at least one of the six affected lots of Focus DAILIES Toric between June 5, 2006 and July 3, 2006:

Lot Number	Pack Size	Label Power	Cylinder	Axis	Base Curve	Diameter
6110160	5-pack	-2.00	-0.75	90	8.6	14.2
6110180	30-pack	-2.00	-0.75	90	8.6	14.2
6110190	30-pack	-2.00	-0.75	90	8.6	14.2
6110210	30-pack	-2.00	-0.75	90	8.6	14.2
6118328	30-pack	-2.00	-0.75	90	8.6	14.2
6118348	5-pack	+0.50	-0.75	180	8.6	14.2

If you have any Focus DAILIES Toric lenses from the affected lot(s), please return them to CIBA Vision in accordance with the instructions on the enclosed business reply form. CIBA Vision will reimburse you for any lenses you return from the affected lots.

We sincerely regret any inconvenience this may cause you and your customers. This recall is being made with the knowledge of the relevant European Competent Authorities. Please feel free to call our Customer Service Department at [Group Company phone] or your CIBA Vision sales representative should you have any questions or concerns.

Yours sincerely,

JULY 2006



Focus DAILIES® Toric Recall 5 and 30-pack

[Insert Group Company ADDRESS]

BUSINESS REPLY FORM

PLEASE COMPLETE THIS FORM AND RETURN WITH PRODUCT OR FAX TO CIBA VISION IMMEDIATELY
[Insert Group Company local FAX NUMBER]

If you are unable to fax or need additional information, please call CIBA Vision Customer Service at
[Insert Group Company local PHONE NUMBER]

Account Name: _____

Account Number: _____ Telephone Number: _____

YES NO I have units of the affected Focus DAILIES® Toric lots (please see below) in stock and will return them to:

 [Insert Group Company local PRODUCT RETURN ADDRESS]

Contact Person (please print): _____

Signature / Date: _____

PRODUCT RETURN INFORMATION

PLEASE COMPLETE THIS SECTION AND RETURN WITH ANY UNITS YOU HAVE IN STOCK OF THE AFFECTED LOTS OF FOCUS DAILIES® TORIC

	Quantity Returned	Date Returned	Returned By
Focus DAILIES® Toric lot #6110160	_____	_____	_____
Focus DAILIES® Toric lot #6110180	_____	_____	_____
Focus DAILIES® Toric lot #6110190	_____	_____	_____
Focus DAILIES® Toric lot #6110210	_____	_____	_____
Focus DAILIES® Toric lot #6118328	_____	_____	_____
Focus DAILIES® Toric lot #6118348	_____	_____	_____

- INSTRUCTIONS**
- Please complete and return the Business Reply Form, even if there is no affected product to return.
 - Please do not include non-recalled products, lots, kits, packs, or competitor's product with your return shipment.
 - Please do not return product through the "normal" CIBA Vision returns process (unless noted). Use the address above.