



URGENT SAFETY INFORMATION

Batch recall of Jelonet (5 cm x 5 cm) "EurimPharm"

To all **wholesalers and pharmacies** supplied with Jelonet (5 cm x 5 cm) by EurimPharm.

(Please also inform your branches and subsidiaries)

Saaldorf-Surheim, 11.06.2021

Dear Sir or Madam,

we would like to inform you about a recall of Jelonet (5 cm x 5 cm) from the manufacturer Smith & Nephew Medical Ltd. which were repackaged and made available on the market by EurimPharm. The following batches are affected:

Product name	Package size	Repack Charge	LOT	Exp.	PZN
Jelonet (5 cm x 5 cm) „EurimPharm“	50 WGA	1000625358	202018	01.04.2025	13946730

Please note that only the listed repacked batches of Jelonet (5 cm x 5 cm) Plaster "EurimPharm" are affected by this letter. All other EurimPharm batches and products are not affected by this safety notice and can continue to be used safely. The repackaging batch can be found on the patch and on the carton.

According to the original manufacturer, there is a potential risk of lack of sterility and leakage of paraffin due to manufacturing defects.

Therefore, we hereby request you to check whether you hold packs of the affected repack batches in stock. According to our distribution documents, we also delivered affected goods to your company in March 2021. **We also ask you to pass on this information to all your customers who may have received the affected batches.**

Please return any remaining packs of the above repacked batches to the following company address for credit (postage costs will be refunded):

**EurimPharm Arzneimittel GmbH
EurimPark 8
83416 Saaldorf-Surheim**

If you have any questions, please contact the EurimPharm Service Hotline on 0800 5000250.



Yours sincerely,
EurimPharm Pharmaceuticals GmbH

